



MISSOURI DEPARTMENT OF HEALTH  
STATE PUBLIC HEALTH LABORATORY  
**DATAMASTER MAINTENANCE REPORT**

BREATH ALCOHOL PROGRAM

JUN -8 2009

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send copy to Department of Health; retain original in department file.

DATAMASTER SN Palmyra Police Department 950264	DATE OF INSPECTION 06-01-2009
LOCATION OF INSTRUMENT (STREET AND CITY) 301 S. Main Palmyra	TIME OF INSPECTION 1218

**CHECKLIST:** Place a check (✓) to the left of each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unchecked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 50 °C	<input checked="" type="checkbox"/> QUARTZ STANDARD
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER
<input checked="" type="checkbox"/> INDICATOR LIGHTS	
<input checked="" type="checkbox"/> TIME AND DATE	
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34 °C ± 0.2°C) 34.0C	
<input checked="" type="checkbox"/> CALIBRATION CHECK - Run three tests using a standard solution. All three tests must be within ± 5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) (USE RECIRCULATION PUMP)	
<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE <input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)	

TEST 1  .099	TEST 2  .098	TEST 3  .099
--------------	--------------	--------------

<input checked="" type="checkbox"/> PERFORM R.F.I. TEST (PRINTOUT ATTACHED)						
<input checked="" type="checkbox"/> NUMBER OF REFUSALS, SINCE LAST MAINTENANCE REPORT, AND NUMBER OF BREATH TESTS IN EACH RANGE AS FOLLOWS: (DO NOT INCLUDE SIMULATOR TESTS)						
REFUSALS 0	(0-.04) 0	(.05-.09) 0	(.10-.14) 1	(.15-.19) 0	(Over .19) 1	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary)

Repro Marketing 0.10 Lot# 08002 Exp 10-13-2010

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT NAME Ronald C. Peer, Jr.
TYPE II PERMIT NUMBER/EXPIRATION DATE 920031/02-18-2011	TELEPHONE NUMBER (573) 769-5540

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.  
LOT NUMBER: 08002  
EXPIRATION DATE: October 13, 2010 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 08002 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed utilizing a gas chromatograph and found to contain 0.1209 gms/dl wt/vol ethyl alcohol in aqueous solution (i.e. ethanol).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of 0.100% +/-3% when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator.

The date of manufacture for this lot number is October 14, 2008. The expiration date for this lot number is October 13, 2010 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President  
RepCo Marking, Inc.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

## BAC DataMaster

Evidence Ticket

STATE OF MISSOURI  
PALMERA POLICE DEPARTMENT

PRO INSTRUMETER SERIAL NUMBER 950264  
06/01/09

ARREST TIME: 12:10  
SUBJECT NAME:  
BLANK TEST  
DOB: 09/09/65 SEX: M  
STATE/D.L.: MO/M087325746  
ARRESTING OFFICER:  
PEER/RON/C  
OFFICER I.D.: 501  
TESTING OFFICER:  
PEER/RON/C  
OFFICER I.D.: 501  
PERMIT NUMBER: 920031  
EXPIRATION DATE: 02/18/11  
MISCELLANEOUS DATA:  
BLANK TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	12:31
INTERNAL STANDARD	VERIFIED	12:31
SUBJECT SAMPLE	.000	12:31
BLANK TEST	.000	12:32

OPERATOR SIGNATURE

*Ronald Clay #501*

Card Stock No.  
60021

REORDER ALL SUPPLIES FROM N.P.A.S.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

## BAC DataMaster

Evidence Ticket

STATE OF MISSOURI  
PALMERA POLICE DEPARTMENT

PRO INSTRUMETER SERIAL NUMBER 950264  
06/01/09

ARREST TIME: 12:10  
SUBJECT NAME:  
RFI TEST  
DOB: 09/09/65 SEX: M  
STATE/D.L.: MO/M087325746  
ARRESTING OFFICER:  
PEER/RON/C  
OFFICER I.D.: 501  
TESTING OFFICER:  
PEER/RON/C  
OFFICER I.D.: 501  
PERMIT NUMBER: 920031  
EXPIRATION DATE: 02/18/11  
MISCELLANEOUS DATA:  
RFI TEST

--- BREATH ANALYSIS ---

BLANK TEST	.000	12:32
INTERNAL STANDARD	VERIFIED	12:32
RADIO INTERFERENCE		

OPERATOR SIGNATURE

*Ronald Clay #501*

Card Stock No.  
60021

REORDER ALL SUPPLIES FROM N.P.A.S.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

## BAC DataMaster

Evidence Ticket

STATE OF MISSOURI  
PALMYRA POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 950264  
06/01/09

### TESTING OFFICER:

PEER/ROM/C

OFFICER I.D.: 501

PERMIT NUMBER: 920031

EXPIRATION DATE: 02/18/11

MISCELLANEOUS DATA:

### --- SUPERVISOR MODE ---

BLANK TEST	.000	12:22
INTERNAL STANDARD	VERIFIED	12:22
EXTERNAL STANDARD	.099	12:22
BLANK TEST	.000	12:23
EXTERNAL STANDARD	.098	12:23
BLANK TEST	.000	12:24
EXTERNAL STANDARD	.099	12:24
BLANK TEST	.000	12:25

N = 3

SIM. = .1

AVG. = .0986

OPERATOR SIGNATURE

*Ronald Clary #501*

Card Stock No.  
60021

REORDER ALL SUPPLIES FROM N.P.A.S.

FACE THIS SIDE DOWN - THIS EDGE IN FIRST

## BAC DataMaster

Evidence Ticket

STATE OF MISSOURI  
PALMYRA POLICE DEPARTMENT

BAC DATA MASTER SERIAL NUMBER 950264  
06/01/09  
12:18

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (24-07-2009): OKAY

HEATERS  
SAMPLE CHAMBER: 50c

FLOW DETECTOR: OKAY

PUMP  
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

### PRINTER TEST

!##%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJK  
LMNOPQRSTUVWXYZ[\]^\_`abcdefg|hijklmnop  
qrstuvwxyz{|}~

ATOR SIGNATURE

*Ronald Clary #501*

ock No.

REORDER ALL SUPPLIES FROM N.P.A.S.



PERMIT  
TYPE II



RONALD C PEER JR.

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 02/18/09

Number 920031

Expires 02/18/2011

MO 590 0771 (7-89)

*John J. Mathewson*

Director of State Public Health Laboratory

*Margaret T. Donnelly*

Director, Department of Health

200, 1 (11, 00)